



**GeneQuery™ Human Noncanonical Wnt Signaling Pathways qPCR Array Kit  
(GQH-WAP)  
Catalog #GK027**

**Product Description**

ScienCell's GeneQuery™ Human Noncanonical Wnt Signaling Pathways qPCR Array Kit (GQH-WAP) is designed to facilitate gene expression profiling of 88 key genes involved in alternative Wnt pathways. This kit emphasizes genes involved in planar cell polarity signaling and calcium signal transduction. Brief examples of how included genes may be grouped according to function are shown below:

- **Planar cell polarity:** DAAM1, CELSR1, FAT1, FZD7, VANGL1
- **Calcium:** CASR, CALM1, FOSL1, FZD4, NFATC1
- **Signaling modulator:** LRP5, LRP6, SPRY1, EFNB, CASR
- **Receptors:** WNT2B, WNT4, WNT5A, WNT9A, WNT11
- **Target genes:** RHOA, RACGAP1, CREBBP, CFL1, FUZ

GeneQuery™ qPCR array kits are qPCR ready in a 96-well plate format, with each well containing one primer set that recognizes and efficiently amplifies a specific target gene's cDNA. The carefully designed primers ensure that: (i) the optimal annealing temperature in qPCR analysis is 65°C (with 2 mM Mg<sup>2+</sup> and no DMSO); (ii) the primer set recognizes all known transcript variants of the target gene, unless otherwise noted; and (iii) only one gene is amplified. Each primer set has been validated by qPCR with melt curve analysis and gel electrophoresis.

**GeneQuery™ qPCR Array Kit Controls**

Each GeneQuery™ plate contains eight controls (Figure 1):

- Five target housekeeping genes (ACTB, GAPDH, LDHA, NONO, and PPIH), which enable normalization of data.
- The Genomic DNA (gDNA) Control (GDC), which detects gDNA contamination in cDNA samples. This primer set targets a non-transcribed region of the genome.
- Positive PCR Control (PPC), which tests whether samples contain inhibitors or other factors that may negatively affect gene expression results. The PPC consists of a predisposed synthetic DNA template and a primer set that can amplify it. The sequence of the DNA template is not present in the human genome and thus tests the efficiency of the polymerase chain reaction itself.
- The No Template Control (NTC), which can be used to monitor DNA contamination introduced during workflow (e.g. from such sources as reagents, tips, and the lab bench).

**Kit Components**

<b>Component</b>	<b>Cat #</b>	<b>Quantity</b>	<b>Storage</b>
GeneQuery™ array plate with lyophilized primers	GK027	1	4°C or -20°C
Optical PCR plate seal	N/A	1	RT
Nuclease-free H <sub>2</sub> O	GQ100-1	2	4°C

**Additional Materials Required (Materials Not Included in Kit)**

<b>Component</b>	<b>Recommended</b>
Reverse transcriptase	First-Strand cDNA Synthesis Master Mix, 4x (ScienCell, Cat #MB6008)
cDNA template	Customers' samples
qPCR master mix	GoldNStart TaqGreen qPCR Master Mix (ScienCell, Cat #MB6018)

**Quality Control**

All primer sets are validated by qPCR with melt curve analysis and analyzed by gel electrophoresis. Single band amplification is confirmed for each set of primers.

**Product Use**

GQH-WAP is for research use only. It is not approved for human or animal use or for application in clinical or *in vitro* diagnostic procedures.

**Shipping and Storage**

This product is shipped at ambient temperature. Upon receipt, the plate should be stored at 4°C and is good for up to 12 months. For long-term storage (>1 year), store at -20°C in a manual defrost freezer.

## Procedures

---

**Note:** The primers in each well are lyophilized.

1. Prior to use, allow plates to warm to room temperature.
2. Briefly centrifuge at 1,500x g for 1 minute before slowly peeling off the seal.
3. Prepare 20  $\mu$ l PCR reactions for one well as shown in Table 1.

Table 1

<b>cDNA template</b>	<b>0.2 – 250 ng</b>
2x qPCR master mix	10 $\mu$ l
Nuclease-free H <sub>2</sub> O	variable
<b>Total volume</b>	<b>20 <math>\mu</math>l</b>

**Important:** *Only use polymerases with hot-start capability to prevent possible primer-dimer formation. Only use nuclease-free reagents in PCR amplification.*

4. Add the mixture of 2x qPCR master mix, cDNA template, and nuclease-free H<sub>2</sub>O to each well containing the lyophilized primers. Seal the plate with the provided optical PCR plate seal.

**Important:** *In NTC control well, do NOT add cDNA template. Add 2x qPCR master mix and nuclease-free H<sub>2</sub>O only.*

5. Briefly centrifuge the plates at 1,500x g for 1 minute at room temperature. For maximum reliability, replicates are strongly recommended (minimum of 3).
6. For PCR program setup, please refer to the instructions of the master mix of the user's choice. We recommend a typical 3-step qPCR protocol for a 200nt amplicon:

### Three-step cycling protocol

Step	Temperature	Time	Number of cycles
Initial denaturation	95°C	10 min	1
Denaturation	95°C	20 sec	40
Annealing	65°C	20 sec	
Extension	72°C	20 sec	
Data acquisition	Plate read		
<i>Recommended</i>	<i>Melting curve analysis</i>		1
Hold	4°C	Indefinite	1

7. (Optional) Load the PCR products on 1.5% agarose gel and perform electrophoresis to confirm the single band amplification in each well.

Figure 1. Layout of GeneQuery™ qPCR array kit controls.




Table 2. Interpretation of control results:

<i>Controls</i>	<i>Results</i>	<i>Interpretation</i>	<i>Suggestions</i>
Housekeeping gene controls	Variability of a housekeeping gene's C <sub>q</sub> value	The expression of the housekeeping gene is variable in samples; cycling program is incorrect	Choose a constantly expressed target, or analyze expression levels of multiple housekeeping genes; use correct cycling program and make sure that all cycle parameters have been correctly entered
gDNA Control (GDC)	C <sub>q</sub> ≥ 35	No gDNA detected	N/A
	C <sub>q</sub> < 35	The sample is contaminated with gDNA	Perform DNase digestion during RNA purification step
Positive PCR Control (PPC)	C <sub>q</sub> > 30; or The C <sub>q</sub> variations > 2 between qPCR Arrays.	Poor PCR performance; possible PCR inhibitor in reactions; cycling program incorrect	Eliminate inhibitor by purifying samples; use correct cycling program and make sure that all cycle parameters have been correctly entered
No Template Control (NTC)	Positive	DNA contamination in workflow	Eliminate sources of DNA contamination (reagents, plastics, etc.)

Rev.1

Figure 2. A typical amplification curve showing the amplification of a qPCR product.





Figure 3. A typical melting peak of a qPCR product.



## **Quantification Method: Comparative $\Delta\Delta Cq$ (Quantification Cycle Value) Method**

---

1. **Note:** Please refer to your qPCR instrument's data analysis software for data analysis. The method provided here serves as guidance for quick manual calculations.

You can use one or more housekeeping genes as a reference to normalize samples.

**Important:** We highly recommend using all 5 housekeeping genes included in this kit: ACTB, GAPDH, LDHA, NONO, and PPIH.

2. For a single housekeeping gene,  $\Delta Cq$  (ref) is the quantification cycle number change for that housekeeping gene (HKG) between an experimental sample and control sample.

$$\Delta Cq \text{ (ref)} = Cq \text{ (HKG, experimental sample)} - Cq \text{ (HKG, control sample)}$$

When using multiple housekeeping genes as a reference, we recommend normalizing using the geometric mean [1] of the expression level change, which is the same as normalizing using the arithmetic mean of  $\Delta Cq$  of the selected housekeeping genes.

$$\Delta Cq \text{ (ref)} = \text{average} (\Delta Cq \text{ (HKG1)}, \Delta Cq \text{ (HKG2)}, \dots, \Delta Cq \text{ (HKG n)}) \text{ (n is the number of housekeeping genes selected)}$$

**If** using all 5 housekeeping genes included in this kit (ACTB, GAPDH, LDHA, NONO, and PPIH) use the following formula:

$$\Delta Cq \text{ (ref)} = (\Delta Cq(\text{ACTB}) + \Delta Cq(\text{GAPDH}) + \Delta Cq(\text{LDHA}) + \Delta Cq(\text{NONO}) + \Delta Cq(\text{PPIH})) / 5$$

**Note:**  $\Delta Cq$  (HKG) =  $Cq$  (HKG, experimental sample) -  $Cq$  (HKG, control sample), and  $\Delta Cq$  (HKG) value can be positive, 0, or negative.

3. For any of your genes of interest (GOI),

$$\Delta Cq \text{ (GOI)} = Cq \text{ (GOI, experimental sample)} - Cq \text{ (GOI, control sample)}$$

$$\Delta\Delta Cq = \Delta Cq \text{ (GOI)} - \Delta Cq \text{ (ref)}$$

$$\text{Normalized GOI expression level fold change} = 2^{-\Delta\Delta Cq}$$

## **References**

[1] Vandesompele J, De Preter K, Pattyn F, Poppe B, Van Roy N, De Paepe A, Speleman F. (2002) "Accurate normalization of real-time quantitative RT-PCR data by geometric averaging of multiple internal control genes." *Genome Biol.* 3(7): 1-12.

**Example: Comparative  $\Delta\Delta Cq$  (Quantification Cycle Value) Method**

Table 3. Cq (Quantification Cycle) values of 2 genes-of-interest and 5 housekeeping genes obtained for experimental and control samples.

Samples	Genes of Interest		Housekeeping Genes				
	GOI1	GOI2	<i>ACTB</i>	<i>GAPDH</i>	<i>LDHA</i>	<i>NONO</i>	<i>PPIH</i>
Experimental	21.61	22.19	17.16	17.84	20.12	19.64	26.40
Control	33.13	26.47	18.20	18.48	20.57	19.50	26.55

$$\begin{aligned}\Delta Cq(\text{ref}) &= (\Delta Cq(\text{ACTB}) + \Delta Cq(\text{GAPDH}) + \Delta Cq(\text{LDHA}) + \Delta Cq(\text{NONO}) + \Delta Cq(\text{PPIH})) / 5 \\ &= ((17.16 - 18.20) + (17.84 - 18.48) + (20.12 - 20.57) + (19.64 - 19.50) + (26.40 - 26.55)) / 5 \\ &= -0.43\end{aligned}$$

$$\begin{aligned}\Delta Cq(\text{GOI1}) &= 21.61 - 33.13 \\ &= -11.52\end{aligned}$$

$$\begin{aligned}\Delta Cq(\text{GOI2}) &= 22.19 - 26.47 \\ &= -4.28\end{aligned}$$

$$\begin{aligned}\Delta\Delta Cq(\text{GOI1}) &= \Delta Cq(\text{GOI1}) - \Delta Cq(\text{ref}) \\ &= -11.52 - (-0.43) \\ &= -11.09\end{aligned}$$

$$\begin{aligned}\Delta\Delta Cq(\text{GOI2}) &= \Delta Cq(\text{GOI2}) - \Delta Cq(\text{ref}) \\ &= -4.28 - (-0.43) \\ &= -3.85\end{aligned}$$

$$\begin{aligned}\text{Normalized GOI1 expression level fold change} &= 2^{-\Delta\Delta Cq(\text{GOI1})} \\ &= 2^{11.09} \\ &= 2180\end{aligned}$$

$$\begin{aligned}\text{Normalized GOI2 expression level fold change} &= 2^{-\Delta\Delta Cq(\text{GOI2})} \\ &= 2^{3.85} \\ &= 14.4\end{aligned}$$

**Conclusion:** Upon treatment, expression level of GOI1 increased 2,180 fold, and expression level of GOI2 increased 14.4 fold.

Rev.1



## GeneQuery™ Human Noncanonical Wnt Signaling Pathways qPCR Array Kit (GQH-WAP)

Catalog #GK027

GeneQuery™ Human Noncanonical Wnt Signaling Pathways qPCR Array Plate Layout\* (*8 controls* in Bold and Italic)

	<b>1</b>	<b>2</b>	<b>3</b>	<b>4</b>	<b>5</b>	<b>6</b>	<b>7</b>	<b>8</b>	<b>9</b>	<b>10</b>	<b>11</b>	<b>12</b>
<b>A</b>	ATN1	CAPN2	CREBBP	DGKE	FAT4	FZD6	ITPR3	PFN1	RACGAP1	SPRY1	WNT4	<b><i>ACTB</i></b>
<b>B</b>	AXIN1	CASR	CREBZF	DVL1	FLNA	FZD7	LIMK1	PLCD1	RGS18	TLR4	WNT5A	<b><i>GAPDH</i></b>
<b>C</b>	AXIN2	CDX2	CSNK1E	DVL2	FOSL1	GSK3B	LRP5	PLCE1	RGS3	VANGL1	WNT5B	<b><i>LDHA</i></b>
<b>D</b>	CALM1	CELSR1	DAAM1	DVL3	FUZ	HGS	LRP6	PLCG1	RHOA	VANGL2	WNT6	<b><i>NONO</i></b>
<b>E</b>	CAMK2A	CELSR2	DCHS1	EFNB1	FZD2	INTU	MAPK8	PRICKLE1	ROR1	WNT11	WNT7A	<b><i>PPIH</i></b>
<b>F</b>	CAMK2B	CELSR3	DCHS2	FAT1	FZD3	INVS	NFATC1	PRICKLE2	ROR2	WNT16	WNT7B	<b><i>GDC</i></b>
<b>G</b>	CAMK4	CFL1	DGKB	FAT2	FZD4	ITPR1	NFATC2	PRKCD	RYK	WNT2B	WNT9A	<b><i>PPC</i></b>
<b>H</b>	CAPN1	CFL2	DGKD	FAT3	FZD5	ITPR2	PCDH8	PTK7	SIAH2	WNT3A	WNT9B	<b><i>NTC</i></b>

\* gene selection may be updated based on new research and development



**Appendix. Plate type choice chart.**

---

**Plate type A**

<b>Brand</b>	<b>Model</b>	<b>kit catalog #</b>
ABI / Life Tech	ABI 5700	GK027-A
	ABI 7000	GK027-A
	ABI 7300	GK027-A
	ABI 7500	GK027-A
	ABI 7700	GK027-A
	ABI 7900 HT	GK027-A
	QuantStudio	GK027-A
	ViiA 7	GK027-A
Bio-Rad	Chromo4	GK027-A
	iCycler	GK027-A
	iQ5	GK027-A
	MyiQ	GK027-A
	MyiQ2	GK027-A
Eppendorf / Life Tech	Matercyler ep realplex 2	GK027-A
	Matercyler ep realplex 4	GK027-A
Stratagene	MX3000P	GK027-A
	MX3005P	GK027-A

**Plate type B**

<b>Brand</b>	<b>Model</b>	<b>kit catalog #</b>
ABI / Life Tech	ABI 7500 Fast	GK027-B
	ABI 7900 HT Fast	GK027-B
	QuantStudio Fast	GK027-B
	StepOnePlus	GK027-B
	ViiA 7 Fast	GK027-B
Bio-Rad	CFX Connect	GK027-B
	CFX96	GK027-B
	DNA Engine Opticon 2	GK027-B
Stratagene	MX4000	GK027-B

**Plate type C**

<b>Brand</b>	<b>Model</b>	<b>kit catalog #</b>
Roche	Lightcycler 96	GK027-C
	Lightcycler 480 (96-well)	GK027-C